

## § 880.1

880.5450 Patient care reverse isolation chamber.  
880.5475 Jet lavage.  
880.5500 AC-powered patient lift.  
880.5510 Non-AC-powered patient lift.  
880.5550 Alternating pressure air flotation mattress.  
880.5560 Temperature regulated water mattress.  
880.5570 Hypodermic single lumen needle.  
880.5580 Acupuncture needle.  
880.5630 Nipple shield.  
880.5640 Lamb feeding nipple.  
880.5680 Pediatric position holder.  
880.5700 Neonatal phototherapy unit.  
880.5725 Infusion pump.  
880.5740 Suction snakebite kit.  
880.5760 Chemical cold pack snakebite kit.  
880.5780 Medical support stocking.  
880.5820 Therapeutic scrotal support.  
880.5860 Piston syringe.  
880.5950 Umbilical occlusion device.  
880.5960 Lice removal kit.  
880.5965 Subcutaneous, implanted, intravascular infusion port and catheter.  
880.5970 Percutaneous, implanted, long-term intravascular catheter.

### Subpart G—General Hospital and Personal Use Miscellaneous Devices

880.6025 Absorbent tipped applicator.  
880.6050 Ice bag.  
880.6060 Medical disposable bedding.  
880.6070 Bed board.  
880.6080 Cardiopulmonary resuscitation board.  
880.6085 Hot/cold water bottle.  
880.6100 Ethylene oxide gas aerator cabinet.  
880.6140 Medical chair and table.  
880.6150 Ultrasonic cleaner for medical instruments.  
880.6175 [Reserved]  
880.6185 Cast cover.  
880.6190 Mattress cover for medical purposes.  
880.6200 Ring cutter.  
880.6230 Tongue depressor.  
880.6250 Patient examination glove.  
880.6265 Examination gown.  
880.6280 Medical insole.  
880.6320 AC-powered medical examination light.  
880.6350 Battery-powered medical examination light.  
880.6375 Patient lubricant.  
880.6430 Liquid medication dispenser.  
880.6450 Skin pressure protectors.  
880.6500 Medical ultraviolet air purifier.  
880.6710 Medical ultraviolet water purifier.  
880.6730 Body waste receptacle.  
880.6740 Vacuum-powered body fluid suction apparatus.  
880.6760 Protective restraint.  
880.6775 Powered patient transfer device.  
880.6785 Manual patient transfer device.  
880.6800 Washers for body waste receptacles.

## 21 CFR Ch. I (4–1–01 Edition)

880.6820 Medical disposable scissors.  
880.6850 Sterilization wrap.  
880.6860 Ethylene oxide gas sterilizer.  
880.6870 Dry-heat sterilizer.  
880.6880 Steam sterilizer.  
880.6885 Liquid chemical sterilants/high level disinfectants.  
880.6890 General purpose disinfectants.  
880.6900 Hand-carried stretcher.  
880.6910 Wheeled stretcher.  
880.6920 Syringe needle introducer.  
880.6960 Irrigating syringe.  
880.6970 Liquid crystal vein locator.  
880.6980 Vein stabilizer.  
880.6990 Infusion stand.

AUTHORITY: 21 U.S.C. 351, 360, 360c, 360e, 360j, 371.

SOURCE: 45 FR 69682–69737, Oct. 21, 1980, unless otherwise noted.

### Subpart A—General Provisions

#### § 880.1 Scope.

(a) This part sets forth the classification of general hospital and personal use devices intended for human use that are in commercial distribution.

(b) The identification of a device in a regulation in this part is not a precise description of every device that is, or will be, subject to the regulation. A manufacturer who submits a pre-market notification submission for a device under part 807 may not show merely that the device is accurately described by the section title and identification provisions of a regulation in this part, but shall state why the device is substantially equivalent to other devices, as required by § 807.87.

(c) To avoid duplicative listings, a general hospital and personal use device that has two or more types of uses (e.g., used both as a diagnostic device and as a therapeutic device) is listed only in one subpart.

(d) References in this part to regulatory sections of the Code of Federal Regulations are to chapter I of title 21, unless otherwise noted.

[52 FR 17738, May 11, 1987]

#### § 880.3 Effective dates of requirement for premarket approval.

A device included in this part that is classified into class III (premarket approval) shall not be commercially distributed after the date shown in the regulation classifying the device unless the manufacturer has an approval