

not intended for administration in humans or in the manufacture of a product.

(iv) You may use human blood or blood components from a donor with a previous record of a reactive screening test(s) for evidence of infection due to a communicable disease agent(s) designated in paragraph (a) of this section, if:

(A) At the time of donation, the donor is shown or was previously shown to be suitable by a requalification method or process found acceptable for such purposes by FDA under § 610.41(b); and

(B) tests performed under paragraphs (a) and (b) of this section are nonreactive.

(v) Anti-HBc reactive donations, otherwise nonreactive when tested as required under this section, may be used for further manufacturing into plasma derivatives without prior FDA approval or a "BIOHAZARD" legend as required under paragraphs (h)(2)(ii)(A) and (h)(2)(ii)(B) of this section.

(vi) You may use human blood or blood components, excluding Source Plasma, that test reactive by a screening test for syphilis as required under paragraph (i) of this section if, consistent with § 640.5 of this chapter, the donation is further tested by an adequate and appropriate test which demonstrates that the reactive screening test is a biological false positive. You must label the blood or blood components with both test results.

(vii) You may use Source Plasma from a donor who tests reactive by a screening test for syphilis as required under § 610.40(i) of this chapter, if the donor meets the requirements of § 640.65(b)(2) of this chapter.

(i) *Syphilis testing.* In addition to the testing otherwise required under this section, you must test by a serological test for syphilis under §§ 640.5(a), 640.14, 640.23(a), 640.33(a), 640.53(a), and 640.65(b)(2) of this chapter.

[66 FR 31162, June 11, 2001]

#### § 610.41 Donor deferral.

(a) You, an establishment that collects human blood or blood components, must defer donors testing reactive by a screening test for evidence of infection due to a communicable dis-

ease agent(s) listed in § 610.40(a) or reactive for a serological test for syphilis under § 610.40(i), from future donations of human blood and blood components, except:

(1) You are not required to defer a donor who tests reactive for anti-HBc or anti-HTLV, types I or II, on only one occasion. When a supplemental (additional, more specific) test for anti-HBc or anti-HTLV, types I and II, has been approved for use under § 610.40(e) by FDA, such a donor must be deferred;

(2) A deferred donor who tests reactive for evidence of infection due to a communicable disease agent(s) listed in § 610.40(a) may serve as a donor for blood or blood components shipped or used under § 610.40(h)(2)(ii);

(3) A deferred donor who showed evidence of infection due to hepatitis B surface antigen (HBsAg) when previously tested under § 610.40(a), (b), and (e) subsequently may donate Source Plasma for use in the preparation of Hepatitis B Immune Globulin (Human) provided the current donation tests nonreactive for HBsAg and the donor is otherwise determined to be suitable;

(4) A deferred donor, who otherwise is determined to be suitable for donation and tests reactive for anti-HBc or for evidence of infection due to HTLV, types I and II, may serve as a donor of Source Plasma;

(5) A deferred donor who tests reactive for a communicable disease agent(s) described under § 610.40(a) or reactive with a serological test for syphilis under § 610.40(i), may serve as an autologous donor under § 610.40(d).

(b) A deferred donor subsequently may be found to be suitable as a donor of blood or blood components by a requalification method or process found acceptable for such purposes by FDA. Such a donor is considered no longer deferred.

[66 FR 31164, June 11, 2001]

#### § 610.42 Restrictions on use for further manufacture of medical devices.

(a) In addition to labeling requirements in subchapter H of this chapter, when a medical device contains human blood or a blood component as a component of the final device, and the human blood or blood component was found to be reactive by a screening test

#### § 610.44

performed under § 610.40(a) and (b) or reactive for syphilis under § 610.40(i), then you must include in the device labeling a statement of warning indicating that the product was manufactured from a donation found to be reactive by a screening test for evidence of infection due to the identified communicable disease agent(s).

(b) FDA may approve an exception or alternative to the statement of warning required in paragraph (a) of this section based on evidence that the reactivity of the human blood or blood component in the medical device presents no significant health risk through use of the medical device.

[66 FR 31164, June 11, 2001]

#### § 610.44 Use of reference panels by manufacturers of test kits.

(a) When available and appropriate to verify acceptable sensitivity and specificity, you, a manufacturer of test kits, must use a reference panel you obtain from FDA or from an FDA designated source to test lots of the following products. You must test each lot of the following products, unless FDA informs you that less frequent testing is appropriate, based on your consistent prior production of products of acceptable sensitivity and specificity:

(1) A test kit approved for use in testing donations of human blood and blood components for evidence of infection due to communicable disease agents listed in § 610.40(a); and

(2) Human immunodeficiency virus (HIV) test kit approved for use in the diagnosis, prognosis, or monitoring of this communicable disease agent.

(b) You must not distribute a lot that is found to be not acceptable for sensitivity and specificity under § 610.44(a). FDA may approve an exception or alternative to this requirement. Applicants must submit such requests in writing. However, in limited circumstances, such requests may be made orally and permission may be given orally by FDA. Oral requests and approvals must be promptly followed by written requests and written approvals.

[66 FR 31164, June 11, 2001]

#### 21 CFR Ch. I (4–1–03 Edition)

#### § 610.46 “Lookback” requirements.

(a) *Quarantine and notification.* (1) All blood and plasma establishments are required to take appropriate action when a donor of Whole Blood, blood components, Source Plasma and Source Leukocytes tests repeatedly reactive for antibody to human immunodeficiency virus (HIV), or otherwise is determined to be unsuitable when tested in accordance with § 610.45. For Whole Blood, blood components, Source Plasma and Source Leukocytes collected from that donor within the 5 years prior to the repeatedly reactive test, if intended for transfusion, or collected within the 6 months prior to the repeatedly reactive test, if intended for further manufacture into injectable products, except those products exempt from quarantine in accordance with § 610.46(c), the blood establishment shall promptly, within 72 hours:

(i) Quarantine all such Whole Blood, blood components, Source Plasma and Source Leukocytes from previous collections held at that establishment; and

(ii) Notify consignees of the repeatedly reactive HIV screening test results so that all Whole Blood, blood components, Source Plasma and Source Leukocytes from previous collections they hold are quarantined.

(2) Consignees notified in accordance with paragraph (a)(1)(ii) of this section shall quarantine Whole Blood, blood components, Source Plasma and Source Leukocytes held at that establishment except as provided in paragraph (c) of this section.

(b) *Further testing and notification of consignees of results.* Blood establishments that have collected Whole Blood, blood components, Source Plasma or Source Leukocytes from a donor as described in paragraph (a) of this section shall perform a licensed, more specific test for HIV on the donor's blood, and in the case of distributed products, further shall notify the consignee(s) of the results of this test, within 30 calendar days after the donor's repeatedly reactive test. Pending the availability of a licensed, more specific test for HIV-2, a second, different screening test for antibody to HIV-2 shall be used along with a licensed, more specific test for HIV-1.