Food and Drug Administration, HHS

purposes, such as selecting or positioning orthopedic implants or guiding the marking of tissue before cutting.

(b) Classification. Class I (general controls). The device is exempt from the premarket notification procedures in subpart E of part 807 of this chapter, subject to the limitations in §888.9.

[52 FR 33702, Sept. 4, 1987, as amended at 66 FR 38815, July 25, 2001]

§ 888.5850 Nonpowered orthopedic traction apparatus and accessories.

- (a) *Identification*. A nonpowered orthopedic traction apparatus is a device that consists of a rigid frame with nonpowered traction accessories, such as cords, pulleys, or weights, and that is intended to apply a therapeutic pulling force to the skeletal system.
- (b) Classification. Class I (general controls). The device is exempt from the premarket notification procedures in subpart E of part 807 of this chapter, subject to the limitations in §888.9. The device is also exempt from the current good manufacturing practice regulations in part 820 of this chapter, with the exception of §820.180, regarding general requirements concerning records, and §820.198, regarding complaint files.

[52 FR 33702, Sept. 4, 1987, as amended at 66 FR 38815, July 25, 2001]

§888.5890 Noninvasive traction component.

- (a) *Identification*. A noninvasive traction component is a device, such as a head halter, pelvic belt, or a traction splint, that does not penetrate the skin and is intended to assist in connecting a patient to a traction apparatus so that a therapeutic pulling force may be applied to the patient's body.
- (b) Classification. Class I (general controls). The device is exempt from the premarket notification procedures in subpart E of part 807 of this chapter, subject to the limitations in §888.9. The device is also exempt from the current good manufacturing practice regulations in part 820 of this chapter, with the exception of §820.180, regarding general requirements concerning

records, and §820.198, regarding complaint files.

[52 FR 33702, Sept. 4, 1987, as amended at 53 FR 52954, Dec. 29, 1988; 66 FR 38815, July 25, 2001]

§888.5940 Cast component.

- (a) *Identification*. A cast component is a device intended for medical purposes to protect or support a cast. This generic type of device includes the cast heel, too cap, cast support, and walking iron
- (b) Classification. Class I (general controls). The device is exempt from the premarket notification procedures in subpart E of part 807 of this chapter, subject to the limitations in §888.9. The device is also exempt from the current good manufacturing practice regulations in part 820 of this chapter, with the exception of §820.180, regarding general requirements concerning records, and §820.198, regarding complaint files.

[52 FR 33702, Sept. 4, 1987, as amended at 53 FR 52954, Dec. 29, 1988; 59 FR 63014, Dec. 7, 1994; 66 FR 38815, July 25, 2001]

§888.5960 Cast removal instrument.

- (a) *Identification*. A cast removal instrument is an AC-powered, hand-held device intended to remove a cast from a patient. This generic type of device includes the electric cast cutter and cast vacuum.
- (b) Classification. Class I (general controls). The device is exempt from the premarket notification procedures in subpart E of part 807 of this chapter, subject to the limitations in §888.9.

[55 FR 48443, Nov. 20, 1990, as amended at 61 FR 1125, Jan. 16, 1996; 66 FR 38816, July 25, 2001]

§888.5980 Manual cast application and removal instrument.

- (a) Identification. A manual cast application and removal instrument is a nonpowered hand-held device intended to be used in applying or removing a cast. This generic type of device includes the cast knife, cast spreader, plaster saw, plaster dispenser, and casting stand.
- (b) Classification. Class I (general controls). The device is exempt from the premarket notification procedures in

21 CFR Ch. I (4-1-03 Edition)

Pt. 890

subpart E of part 807 of this chapter, subject to the limitations in §888.9. The device is also exempt from the current good manufacturing practice regulations in part 820 of this chapter, with the exception of §820.180, regarding general requirements concerning records, and §820.198, regarding complaint files.

[52 FR 33702, Sept. 4, 1987, as amended at 53 FR 52954, Dec. 29, 1988; 66 FR 38816, July 25,

PART 890—PHYSICAL MEDICINE **DEVICES**

Subpart A—General Provisions

Sec.

890.1 Scope.

890.3 Effective dates of requirement for premarket approval.

890.9 Limitations of exemptions from section 510(k) of the Federal Food, Drug, and Cosmetic Act (the act).

Subpart B—Physical Medicine Diagnostic **Devices**

890.1175 Electrode cable.

890.1225 Chronaximeter.

890.1375 Diagnostic electromyograph.

890.1385 Diagnostic electromyograph needle electrode.

890.1450 Powered reflex hammer.

890.1575 Force-measuring platform.

890.1600 Intermittent pressure measurement system.

890.1615 Miniature pressure transducer.

890.1850 Diagnostic muscle stimulator.

890.1925 Isokinetic testing and evaluation system.

Subpart C [Reserved]

Subpart D-Physical Medicine Prosthetic Devices

890.3025 Prosthetic and orthotic accessory.

890.3075 Cane.

890.3100 Mechanical chair.

890.3110 Electric positioning chair.

890.3150 Crutch.

890.3175 Flotation cushion.

External limb orthotic component. 890.3410

890.3420 External limb prosthetic component.

890.3475 Limb orthosis.

890 3490 Truncal orthosis.

890.3500 External assembled lower limb prosthesis.

890.3520 Plinth

890.3610 Rigid pneumatic structure orthosis.

890.3640 Arm sling.

890.3665 Congenital hip dislocation abduction splint.

890.3675 Denis Brown splint

890.3690 Powered wheeled stretcher.

890.3700 Nonpowered communication system.

890.3710 Powered communication system.

890.3725 Powered environmental control system.

890.3750 Mechanical table.

890.3760 Powered table.

Cane, crutch, and walker tips and 890.3790 pads.

890.3800 Motorized three-wheeled vehicle.

890.3825 Mechanical walker.

890.3850 Mechanical wheelchair. 890.3860 Powered wheelchair.

Special grade wheelchair. 890.3880

890.3890 Stair-climbing wheelchair.

890.3900 Standup wheelchair.

890.3910 Wheelchair accessory

890.3920 Wheelchair component.

890.3930 Wheelchair elevator.

890 3940 Wheelchair platform scale.

Subpart E [Reserved]

Subpart F—Physical Medicine Therapeutic **Devices**

890.5050 Daily activity assist device.

890.5100 Immersion hydrobath.

890.5110 Paraffin bath.

890.5125 Nonpowered sitz bath.

890.5150 Powered patient transport. 890.5160 Air-fluidized bed.

890.5170 Powered flotation therapy bed.

890.5180 Manual patient rotation bed.

890.5225 Powered patient rotation bed.

890.5250 Moist steam cabinet.

890.5275 Microwave diathermy. 890.5290 Shortwave diathermy.

890.5300 Ultrasonic diathermy.

890.5350

Exercise component.

890.5360 Measuring exercise equipment. 890.5370

Nonmeasuring exercise equipment. 890.5380 Powered exercise equipment.

890.5410 Powered finger exerciser.

890.5500 Infrared lamp.

890.5525 Iontophoresis device.

890.5575 Powered external limb overload warning device.

890.5650 Powered inflatable tube massager.

890.5660 Therapeutic massager.

890.5700 Cold pack.

890.5710 Hot or cold disposable pack. 890.5720 Water circulating hot or cold pack.

890.5730 Moist heat pack.

890.5740 Powered heating pad.

890.5765 Pressure-applying device.

890.5850 Powered muscle stimulator.

890.5860 Ultrasound and muscle stimulator.

890.5880 Multi-function physical therapy table.

890.5900 Powered traction equipment.

890.5925 Traction accessory.

890.5940 Chilling unit.