

§ 53.2 General requirements for a reference method determination.

The following general requirements for a reference method determination are summarized in table A-1 of this subpart.

(a) *Manual methods.* (1) For measuring sulfur dioxide (SO₂) and lead, Appendices A and G of part 50 of this chapter specify unique manual reference methods for those pollutants. Except as provided in § 53.16, other manual methods for SO₂ and lead will not be considered for reference method determinations under this part.

(2) A reference method for measuring PM₁₀ must be a manual method that meets all requirements specified in appendix J of part 50 of this chapter and must include a PM₁₀ sampler that has been shown in accordance with this part to meet all requirements specified in subparts A and D of this part.

(3) A reference method for measuring PM_{2.5} must be a manual method that meets all requirements specified in appendix L of part 50 of this chapter and must include a PM_{2.5} sampler that has been shown in accordance with this part to meet the applicable requirements specified in subparts A and E of this part. Further, reference method samplers must be manufactured in an ISO 9001-registered facility, as defined in § 53.1 and as set forth in § 53.51, and the Product Manufacturing Checklist set forth in subpart E of this part must be completed by an ISO-certified auditor, as defined in § 53.1, and submitted to EPA annually to retain a PM_{2.5} reference method designation.

(b) *Automated methods.* An automated reference method for measuring carbon monoxide (CO), ozone (O₃), and nitrogen dioxide (NO₂) must utilize the measurement principle and calibration procedure specified in the appropriate appendix to part 50 of this chapter and must have been shown in accordance with this part to meet the requirements specified in subpart B of this part.

§ 53.3 General requirements for an equivalent method determination.

(a) *Manual methods.* A manual equivalent method must have been shown in accordance with this part to satisfy the applicable requirements specified in

subpart C of this part. In addition, PM₁₀ or PM_{2.5} samplers associated with manual equivalent methods for PM₁₀ or PM_{2.5} must have been shown in accordance with this part to satisfy the following additional requirements:

(1) A PM₁₀ sampler associated with a manual method for PM₁₀ must satisfy the requirements of subpart D of this part.

(2) A PM_{2.5} Class I equivalent method sampler must satisfy all requirements of subparts C and E of this part, which include appropriate demonstration that each and every deviation or modification from the reference method sampler specifications does not significantly alter the performance of the sampler.

(3) A PM_{2.5} Class II equivalent method sampler must satisfy the applicable requirements of subparts C, E, and F of this part.

(4) Requirements for PM_{2.5} Class III equivalent method samplers are not provided in this part because of the wide range of non-filter-based measurement technologies that could be applied and the likelihood that these requirements will have to be specifically adapted for each such type of technology. Specific requirements will be developed as needed and may include selected requirements from subparts C, E, or F of this part or other requirements not contained in this part.

(5) All designated equivalent methods for PM_{2.5} must be manufactured in an ISO 9001-registered facility, as defined in § 53.1 and as set forth in § 53.51, and the Product Manufacturing Checklist set forth in subpart E of this part must be completed by an ISO-certified auditor, as defined in § 53.1, and submitted to EPA annually to retain a PM_{2.5} equivalent method designation.

(b) *Automated methods.* (1) Automated equivalent methods for pollutants other than PM_{2.5} or PM₁₀ must have been shown in accordance with this part to satisfy the requirements specified in subparts B and C of this part.

(2) Automated equivalent methods for PM₁₀ must have been shown in accordance with this part to satisfy the requirements of subparts C and D of this part.

(3) Requirements for PM_{2.5} Class III automated equivalent methods for