

(1) Precautions to be followed to protect the patient from unnecessary exposure to radiation;

(2) Precautions to be followed to protect an individual supporting the patient during X-ray procedures from unnecessary exposure to radiation;

(3) Precautions to be followed to protect other individuals in the surrounding environment from exposure to radiation;

(4) Precautions to be followed to protect the operator of portable X-ray equipment from unnecessary exposure to radiation;

(5) Considerations in determining the area which will receive the primary beam;

(6) Determination of the time interval at which to check personnel radiation monitors;

(7) Use of the personnel radiation monitor in providing an additional check on safety of equipment;

(8) Proper use and maintenance of equipment;

(9) Proper maintenance of records;

(10) Technical problems which may arise and methods of solution;

(11) Protection against electrical hazards;

(12) Hazards of excessive exposure to radiation.

(c) *Standard: Employee records.* Records are maintained and include evidence that—

(1) Each employee is qualified for his or her position by means of training and experience; and

(2) Employees receive adequate health supervision.

[34 FR 388, Jan. 10, 1969. Redesignated at 42 FR 52826, Sept. 30, 1977, and amended at 53 FR 12015, Apr. 12, 1988; 60 FR 45086, Aug. 30, 1995]

§ 486.106 Condition for coverage: Referral for service and preservation of records.

All portable X-ray services performed for Medicare beneficiaries are ordered by a doctor of medicine or doctor of osteopathy and records are properly preserved.

(a) *Standard—referral by a physician.* Portable X-ray examinations are performed only on the order of a doctor of medicine or doctor of osteopathy li-

censed to practice in the State. The supplier's records show that:

(1) The X-ray test was ordered by a licensed doctor of medicine or doctor of osteopathy, and

(2) Such physician's written, signed order specifies the reason an X-ray test is required, the area of the body to be exposed, the number of radiographs to be obtained, and the views needed; it also includes a statement concerning the condition of the patient which indicates why portable X-ray services are necessary.

(b) *Standard—records of examinations performed.* The supplier makes for each patient a record of the date of the X-ray examination, the name of the patient, a description of the procedures ordered and performed, the referring physician, the operator(s) of the portable X-ray equipment who performed the examination, the physician to whom the radiograph was sent, and the date it was sent.

(c) *Standard—preservation of records.* Such reports are maintained for a period of at least 2 years, or for the period of time required by State law for such records (as distinguished from requirements as to the radiograph itself), whichever is longer.

[34 FR 388, Jan. 10, 1969. Redesignated at 42 FR 52826, Sept. 30, 1977. Further redesignated and amended at 60 FR 2326, Jan. 9, 1995; 60 FR 45086, Aug. 30, 1995]

§ 486.108 Condition for coverage: Safety standards.

X-ray examinations are conducted through the use of equipment which is free of unnecessary hazards for patients, personnel, and other persons in the immediate environment, and through operating procedures which provide minimum radiation exposure to patients, personnel, and other persons in the immediate environment.

(a) *Standard—tube housing and devices to restrict the useful beam.* The tube housing is of diagnostic type. Diaphragms, cones, or adjustable collimators capable of restricting the useful beam to the area of clinical interest are used and provide the same degree of protection as is required of the housing.