

§ 352.71 Light source (solar simulator).

A solar simulator used for determining the SPF of a sunscreen drug product should be filtered so that it provides a continuous emission spectrum from 290 to 400 nanometers similar to sunlight at sea level from the sun at a zenith angle of 10° it has less than 1 percent of its total energy output contributed by nonsolar wavelengths shorter than 290 nanometers; and it has not more than 5 percent of its total energy output contributed by wavelengths longer than 400 nanometers. In addition, a solar simulator should have no significant time-related fluctuations in radiation emissions after an appropriate warmup time, and it should have good beam uniformity (within 10 percent) in the exposure plane. To ensure that the solar simulator delivers the appropriate spectrum of UV radiation, it must be measured periodically with an accurately-calibrated spectroradiometer system or equivalent instrument.

§ 352.72 General testing procedures.

(a) *Selection of test subjects (male and female).* (1) Only fair-skin subjects with skin types I, II, and III using the following guidelines shall be selected:

Selection of Fair-skin Subjects

Skin Type and Sunburn and Tanning History (Based on first 30 to 45 minutes sun exposure after a winter season of no sun exposure.)

- I—Always burns easily; never tans (sensitive).
- II—Always burns easily; tans minimally (sensitive).
- III—Burns moderately; tans gradually (light brown) (normal).
- IV—Burns minimally; always tans well (moderate brown) (normal).
- V—Rarely burns; tans profusely (dark brown) (insensitive).
- VI—Never burns; deeply pigmented (insensitive).

(2) A medical history shall be obtained from all subjects with emphasis on the effects of sunlight on their skin. Ascertain the general health of the individual, the individual's skin type (I, II, or III), whether the individual is taking medication (topical or systemic) that is known to produce abnormal sunlight responses, and whether the individual is subject to any abnormal

responses to sunlight, such as a phototoxic or photoallergic response.

(b) *Test site inspection.* The physical examination shall determine the presence of sunburn, suntan, scars, active dermal lesions, and uneven skin tones on the areas of the back to be tested. The presence of nevi, blemishes, or moles will be acceptable if in the physician's judgment they will not interfere with the study results. Excess hair on the back is acceptable if the hair is clipped or shaved.

(c) *Informed consent.* Legally effective written informed consent must be obtained from all individuals.

(d) *Test site delineation—(1) Test site area.* A test site area serves as an area for determining the subject's MED after application of either the sunscreen standard or the test sunscreen product, or for determining the subject's MED when the skin is unprotected (control site). The area to be tested shall be the back between the beltline and the shoulder blade (scapulae) and lateral to the midline. Each test site area for applying a product or the standard sunscreen shall be a minimum of 50-square centimeters, e.g., 5 × 10 centimeters. The test site areas are outlined with ink. If the person is to be tested in an upright position, the lines shall be drawn on the skin with the subject upright. If the subject is to be tested while prone, the markings shall be made with the subject prone.

(2) *Test subsite area.* Each test site area shall be divided into at least three test subsite areas that are at least 1 square centimeter. Usually four or five subsites are employed. Each test subsite within a test site area is subjected to a specified dosage of UV radiation, in a series of UV radiation exposures, in which the test site area is exposed for the determination of the MED.

(e) *Application of test materials.* To ensure standardized reporting and to define a product's SPF value, the application of the product shall be expressed on a weight basis per unit area which establishes a standard film. Both the test sunscreen product and the standard sunscreen application shall be 2 milligrams per square centimeter. For oils and most lotions, the viscosity is