

standard specifications of the American Society for Testing and Materials (ASTM) for such instruments. Such specific gravity hydrometers shall be of a precision grade, standardization temperature 60 °/60 °F., and provided in the following ranges and subdivisions:

Range	Subdivision
1.0000 to 1.0500	0.0005
1.0500 to 1.1000	0.0005
1.1000 to 1.1500	0.0005
1.1500 to 1.2000	0.0005
1.2000 to 1.2500	0.0005

No instrument shall be in error by more than 0.0005 specific gravity.

(b) A certificate of accuracy prepared by the instrument manufacturer for the instrument shall be furnished to the appropriate ATF officer.

(c) *Incorporation by reference.* The “Standard Specification for ASTM Hydrometers,” (E 100–72 (1978)), published in the “1980 Annual Book of ASTM Standards” (STP 25 1062 (1980)), is incorporated by reference in this part. This incorporation by reference was approved by the Director of the Federal Register on March 23, 1981. This publication may be inspected at the National Archives and Records Administration (NARA), and is available from the American Society for Testing and Materials, 1916 Race Street, Philadelphia, Pennsylvania 19103. For information on the availability of this material at NARA, call 202-741-6030, or go to: http://www.archives.gov/federal_register/code_of_federal_regulations/ibr_locations.html.

(Sec. 201, Pub. L. 85-859, 72 Stat. 1358, as amended (26 U.S.C. 5204); 80 Stat. 383, as amended (5 U.S.C. 552(a)))

[T.D. ATF-198, 50 FR 8535, Mar. 1, 1985, as amended by T.D. ATF-381, 61 FR 37004, July 16, 1996; 69 FR 18803, Apr. 9, 2004]

§ 30.25 Use of precision specific gravity hydrometers.

The provisions of § 30.23 respecting the care, handling, and use of precision instruments shall be followed with respect to the care, handling, and use of precision grade specific gravity hydrometers. Specific gravity hydrometers shall be read to the nearest subdivision. Because of temperature den-

sity relationships and the selection of the standardization temperature of 60 °/60 °F., the specific gravity readings will be greater at temperatures below 60 degrees Fahrenheit and less at temperatures above 60 degrees Fahrenheit. Hence, correction of the specific gravity readings will be made for temperature other than 60 degrees Fahrenheit. Such correction may be ascertained by dividing the specific gravity hydrometer reading by the applicable correction factor in Table 7.

Example: The specific gravity hydrometer reading is 1.1525, the thermometer reading is 68 degrees Fahrenheit, and the true proof of the spirits is 115 degrees. The correct specific gravity reading will be ascertained as follows:

(a) From Table 7, the correction factor for 115° proof at 68 °F. is 0.996.

(b) 1.1525 divided by 0.996=1.1571, the corrected specific gravity.

(Sec. 201, Pub. L. 85-859, 72 Stat. 1358, as amended (26 U.S.C. 5204))

Subpart D—Gauging Procedures

§ 30.31 Determination of proof.

(a) *General.* The proof of spirits shall be determined to the nearest tenth degree which shall be the proof used in determining the proof gallons.

(b) *Solids content not more than 600 milligrams.* Except as otherwise authorized by the appropriate ATF officer, the proof of spirits containing not more than 600 milligrams of solids per 100 milliliters of spirits shall be determined by the use of a hydrometer and thermometer in accordance with the provisions of § 30.23 except that if such spirits contain solids in excess of 400 milligrams but not in excess of 600 milligrams per 100 milliliters at gauge proof, there shall be added to the proof so determined the obscuration determined as prescribed in § 30.32.

(c) *Solids content over 600 milligrams.* If such spirits contain solids in excess of 600 milligrams per 100 milliliters at gauge proof, the proof shall be determined on the basis of true proof determined as follows:

(1) By the use of a hydrometer and a thermometer after the spirits have been distilled in a small laboratory