

TABLE 1 OF § 1065.303—SUMMARY OF REQUIRED CALIBRATION AND VERIFICATIONS—Continued

| Type of calibration or verification | Minimum frequency ^a |
|--|---|
| § 1065.350: CO ₂ NDIR H ₂ O interference | Upon initial installation and after major maintenance. |
| § 1065.355: CO NDIR CO ₂ and H ₂ O interference. | Upon initial installation and after major maintenance. |
| § 1065.360: FID optimization, etc. | <i>Calibrate, optimize, and determine CH₄ response:</i> upon initial installation and after major maintenance. <i>Verify CH₄ response:</i> upon initial installation, within 185 days before testing, and after major maintenance. |
| § 1065.362: raw exhaust FID O ₂ interference. | Upon initial installation, after FID optimization according to § 1065.360, and after major maintenance. |
| § 1065.365: nonmethane cutter penetration | Upon initial installation, within 185 days before testing, and after major maintenance. |
| § 1065.370: CLD CO ₂ and H ₂ O quench | Upon initial installation and after major maintenance. |
| § 1065.372: NDUV HC and H ₂ O interference. | Upon initial installation and after major maintenance. |
| § 1065.376: chiller NO ₂ penetration | Upon initial installation and after major maintenance. |
| § 1065.378: NO ₂ -to-NO converter conversion. | Upon initial installation, within 35 days before testing, and after major maintenance. |
| § 1065.390: PM balance and weighing | <i>Independent verification:</i> upon initial installation, within 370 days before testing, and after major maintenance. <i>Zero, span, and reference sample verifications:</i> within 12 hours of weighing, and after major maintenance. |
| § 1065.395: Inertial PM balance and weighing. | <i>Independent verification:</i> upon initial installation, within 370 days before testing, and after major maintenance. <i>Other verifications:</i> upon initial installation and after major maintenance. |

^a Perform calibrations and verifications more frequently, according to measurement system manufacturer instructions and good engineering judgment.

§ 1065.305 Verifications for accuracy, repeatability, and noise.

(a) This section describes how to determine the accuracy, repeatability, and noise of an instrument. Table 1 of § 1065.205 specifies recommended values for individual instruments.

(b) We do not require you to verify instrument accuracy, repeatability, or noise.

However, it may be useful to consider these verifications to define a specification for a new instrument, to verify the performance of a new instrument upon delivery, or to troubleshoot an existing instrument.

(c) In this section we use the letter “y” to denote a generic measured quantity, the superscript over-bar to denote an arithmetic mean (such as \bar{y}), and the subscript “*ref*” to denote the reference quantity being measured.

(d) Conduct these verifications as follows:

(1) Prepare an instrument so it operates at its specified temperatures, pressures, and flows. Perform any instrument linearization or calibration procedures prescribed by the instrument manufacturer.

(2) Zero the instrument as you would before an emission test by introducing a zero signal. Depending on the instru-

ment, this may be a zero-concentration gas, a reference signal, a set of reference thermodynamic conditions, or some combination of these. For gas analyzers, use a zero gas that meets the specifications of § 1065.750.

(3) Span the instrument as you would before an emission test by introducing a span signal. Depending on the instrument, this may be a span-concentration gas, a reference signal, a set of reference thermodynamic conditions, or some combination of these. For gas analyzers, use a span gas that meets the specifications of § 1065.750.

(4) Use the instrument to quantify a NIST-traceable reference quantity, y_{ref} . For gas analyzers the reference gas must meet the specifications of § 1065.750. Select a reference quantity near the mean value expected during testing. For all gas analyzers, use a quantity near the flow-weighted mean concentration expected at the standard or expected during testing, whichever is greater. For a noise verification, use the same zero gas from paragraph (e) of this section as the reference quantity. In all cases, allow time for the instrument to stabilize while it measures the reference quantity. Stabilization time

may include time to purge an instrument and time to account for its response.

(5) Sample and record values for 30 seconds, record the arithmetic mean, \bar{y}_i , and record the standard deviation, σ_i , of the recorded values. Refer to §1065.602 for an example of calculating arithmetic mean and standard deviation.

(6) Also, if the reference quantity is not absolutely constant, which might be the case with a reference flow, sample and record values of y_{ref} for 30 seconds and record the arithmetic mean of the values, \bar{y}_{ref} . Refer to §1065.602 for an example of calculating arithmetic mean.

(7) Subtract the reference value, y_{ref} (or \bar{y}_{ref}), from the arithmetic mean, \bar{y}_i . Record this value as the error, ϵ_i .

(8) Repeat the steps specified in paragraphs (d)(2) through (6) of this section until you have ten arithmetic means ($\bar{y}_1, \bar{y}_2, \bar{y}_i, \dots, \bar{y}_{10}$), ten standard deviations ($\sigma_1, \sigma_2, \sigma_i, \dots, \sigma_{10}$), and ten errors ($\epsilon_1, \epsilon_2, \epsilon_i, \dots, \epsilon_{10}$).

(9) Use the following values to quantify your measurements:

(i) *Accuracy.* Instrument accuracy is the absolute difference between the reference quantity, y_{ref} (or \bar{y}_{ref}), and the arithmetic mean of the ten \bar{y}_i, \bar{y} values. Refer to the example of an accuracy calculation in §1065.602. We recommend that instrument accuracy be within the specifications in Table 1 of §1065.205.

(ii) *Repeatability.* Repeatability is two times the standard deviation of the ten errors (that is, repeatability = $2 \cdot \sigma_\epsilon$). Refer to the example of a standard-deviation calculation in §1065.602. We recommend that instrument repeatability be within the specifications in Table 1 of §1065.205.

(iii) *Noise.* Noise is two times the root-mean-square of the ten standard deviations (that is, noise = $2 \cdot rms_\sigma$) when the reference signal is a zero-quantity signal. Refer to the example of a root-mean-square calculation in §1065.602. We recommend that instrument noise be within the specifications in Table 1 of §1065.205. Use this value in the noise correction specified in §1065.657.

(10) You may use a measurement instrument that does not meet the accu-

racy, repeatability, or noise specifications in Table 1 of §1065.205, as long as you meet the following criteria:

(i) Your measurement systems meet all the other required calibration, verification, and validation specifications in subparts D, F, and J of this part, as applicable.

(ii) The measurement deficiency does not adversely affect your ability to demonstrate compliance with the applicable standards.

§ 1065.307 Linearity verification.

(a) *Scope and frequency.* Perform a linearity verification on each measurement system listed in Table 1 of this section at least as frequently as indicated in the table, consistent with measurement system manufacturer recommendations and good engineering judgment. Note that this linearity verification may replace requirements we previously referred to as “calibrations”. The intent of a linearity verification is to determine that a measurement system responds proportionally over the measurement range of interest. A linearity verification generally consists of introducing a series of at least 10 reference values to a measurement system. The measurement system quantifies each reference value. The measured values are then collectively compared to the reference values by using a least squares linear regression and the linearity criteria specified in Table 1 of this section.

(b) *Performance requirements.* If a measurement system does not meet the applicable linearity criteria in Table 1 of this section, correct the deficiency by re-calibrating, servicing, or replacing components as needed. Before you may use a measurement system that does not meet linearity criteria, you must demonstrate to us that the deficiency does not adversely affect your ability to demonstrate compliance with the applicable standards.

(c) *Procedure.* Use the following linearity verification protocol, or use good engineering judgment to develop a different protocol that satisfies the intent of this section, as described in paragraph (a) of this section:

(1) In this paragraph (c), we use the letter “y” to denote a generic measured quantity, the superscript over-bar