

**§ 493.1239**

(c) At least twice annually, the laboratory must verify the accuracy of the following:

(1) Any test or procedure it performs that is not included in subpart I of this part.

(2) Any test or procedure listed in subpart I of this part for which compatible proficiency testing samples are not offered by a CMS-approved proficiency testing program.

(d) All proficiency testing evaluation and verification activities must be documented.

**§ 493.1239 Standard: General laboratory systems quality assessment.**

(a) The laboratory must establish and follow written policies and procedures for an ongoing mechanism to monitor, assess, and, when indicated, correct problems identified in the general laboratory systems requirements specified at §§ 493.1231 through 493.1236.

(b) The general laboratory systems quality assessment must include a review of the effectiveness of corrective actions taken to resolve problems, revision of policies and procedures necessary to prevent recurrence of problems, and discussion of general laboratory systems quality assessment reviews with appropriate staff.

(c) The laboratory must document all general laboratory systems quality assessment activities.

[68 FR 3703, Jan. 24, 2003; 68 FR 50724, Aug. 22, 2003]

PREANALYTIC SYSTEMS

**§ 493.1240 Condition: Preanalytic systems.**

Each laboratory that performs non-waived testing must meet the applicable preanalytic system(s) requirements in §§ 493.1241 and 493.1242, unless HHS approves a procedure, specified in Appendix C of the State Operations Manual (CMS Pub. 7), that provides equivalent quality testing. The laboratory must monitor and evaluate the overall quality of the preanalytic systems and correct identified problems as specified in § 493.1249 for each specialty and subspecialty of testing performed.

**42 CFR Ch. IV (10–1–06 Edition)**

**§ 493.1241 Standard: Test request.**

(a) The laboratory must have a written or electronic request for patient testing from an authorized person.

(b) The laboratory may accept oral requests for laboratory tests if it solicits a written or electronic authorization within 30 days of the oral request and maintains the authorization or documentation of its efforts to obtain the authorization.

(c) The laboratory must ensure the test requisition solicits the following information:

(1) The name and address or other suitable identifiers of the authorized person requesting the test and, if appropriate, the individual responsible for using the test results, or the name and address of the laboratory submitting the specimen, including, as applicable, a contact person to enable the reporting of imminently life threatening laboratory results or panic or alert values.

(2) The patient's name or unique patient identifier.

(3) The sex and age or date of birth of the patient.

(4) The test(s) to be performed.

(5) The source of the specimen, when appropriate.

(6) The date and, if appropriate, time of specimen collection.

(7) For Pap smears, the patient's last menstrual period, and indication of whether the patient had a previous abnormal report, treatment, or biopsy.

(8) Any additional information relevant and necessary for a specific test to ensure accurate and timely testing and reporting of results, including interpretation, if applicable.

(d) The patient's chart or medical record may be used as the test requisition or authorization but must be available to the laboratory at the time of testing and available to CMS or a CMS agent upon request.

(e) If the laboratory transcribes or enters test requisition or authorization information into a record system or a laboratory information system, the laboratory must ensure the information is transcribed or entered accurately.