

(4) Make a reasonable attempt to have available monthly serum specimens for all potential transplant recipients for periodic antibody screening and crossmatch.

(5) Have available and follow a written policy consistent with clinical transplant protocols for the frequency of screening potential transplant recipient sera for preformed HLA-specific antibodies.

(6) Check each antibody screening by testing, at a minimum the following:

(i) A positive control material containing antibodies of the appropriate isotype for the assay.

(ii) A negative control material.

(7) As applicable, have available and follow written criteria and procedures for antibody identification to the level appropriate to support clinical transplant protocol.

(e) *Crossmatching.* The laboratory must do the following:

(1) Use a technique(s) documented to have increased sensitivity in comparison with the basic complement-dependent microlymphocytotoxicity assay.

(2) Have available and follow written criteria for the following:

(i) Selecting appropriate patient serum samples for crossmatching.

(ii) The preparation of donor cells or cellular extracts (for example, solubilized antigens and nucleic acids), as applicable to the crossmatch technique(s) performed.

(3) Check each crossmatch and compatibility test for HLA Class II antigenic differences using control materials to monitor the test components and each phase of the test system to ensure acceptable performance.

(f) *Transplantation.* Laboratories performing histocompatibility testing for transfusion and transplantation purposes must do the following:

(1) Have available and follow written policies and protocols specifying the histocompatibility testing (that is, HLA typing, antibody screening, compatibility testing and crossmatching) to be performed for each type of cell, tissue or organ to be transfused or transplanted. The laboratory's policies must include, as applicable—

(i) Testing protocols for cadaver donor, living, living-related, and combined organ and tissue transplants;

(ii) Testing protocols for patients at high risk for allograft rejection; and

(iii) The level of testing required to support clinical transplant protocols (for example, antigen or allele level).

(2) For renal allotransplantation and combined organ and tissue transplants in which a kidney is to be transplanted, have available results of final crossmatches before the kidney is transplanted.

(3) For nonrenal transplantation, if HLA testing and final crossmatches were not performed prospectively because of an emergency situation, the laboratory must document the circumstances, if known, under which the emergency transplant was performed, and records of the transplant must reflect any information provided to the laboratory by the patient's physician.

(g) *Documentation.* The laboratory must document all control procedures performed, as specified in this section.

[68 FR 3703, Jan. 24, 2003; 68 FR 50724, Aug. 22, 2003]

§ 493.1281 Standard: Comparison of test results.

(a) If a laboratory performs the same test using different methodologies or instruments, or performs the same test at multiple testing sites, the laboratory must have a system that twice a year evaluates and defines the relationship between test results using the different methodologies, instruments, or testing sites.

(b) The laboratory must have a system to identify and assess patient test results that appear inconsistent with the following relevant criteria, when available:

(1) Patient age.

(2) Sex.

(3) Diagnosis or pertinent clinical data.

(4) Distribution of patient test results.

(5) Relationship with other test parameters.

(c) The laboratory must document all test result comparison activities.