

**§ 606.65**

(c) Equipment employed in the sterilization of materials used in blood collection or for disposition of contaminated products shall be designed, maintained and utilized to ensure the destruction of contaminating microorganisms. The effectiveness of the sterilization procedure shall be no less than that achieved by an attained temperature of 121.5 °C (251 °F) maintained for 20 minutes by saturated steam or by an attained temperature of 170 °C (338 °F) maintained for 2 hours with dry heat.

[40 FR 53532, Nov. 18, 1975; 40 FR 55849, Dec. 2, 1975, as amended at 45 FR 9261, Feb. 12, 1980; 57 FR 11263, Apr. 2, 1992; 57 FR 12862, Apr. 13, 1992]

**§ 606.65 Supplies and reagents.**

All supplies and reagents used in the collection, processing, compatibility testing, storage and distribution of blood and blood components shall be stored in a safe, sanitary and orderly manner.

(a) All surfaces coming in contact with blood and blood components intended for transfusion shall be sterile, pyrogen-free, and shall not interact with the product in such a manner as to have an adverse effect upon the safety, purity, potency or effectiveness of the product. All final containers and closures for blood and blood components not intended for transfusion shall be clean and free of surface solids and other contaminants.

(b) Each blood collecting container and its satellite container(s), if any, shall be examined visually for damage or evidence of contamination prior to its use and immediately after filling. Such examination shall include inspection for breakage of seals, when indicated, and abnormal discoloration. Where any defect is observed, the container shall not be used, or, if detected after filling, shall be properly discarded.

(c) Representative samples of each lot of the following reagents or solutions shall be tested on a regularly scheduled basis by methods described in the Standard Operating Procedures Manual to determine their capacity to perform as required:

Reagent or solution	Frequency of testing
Anti-human globulin .....	Each day of use.

**21 CFR Ch. I (4-1-07 Edition)**

Reagent or solution	Frequency of testing
Blood grouping reagents .....	Do.
Lectins .....	Do.
Antibody screening and reverse grouping cells.	Do.
Hepatitis test reagents .....	Each run.
Syphilis serology reagents ....	Do.
Enzymes .....	Each day of use.

(d) Supplies and reagents that do not bear an expiration date shall be stored in such a manner that the oldest is used first.

(e) Supplies and reagents shall be used in a manner consistent with instructions provided by the manufacturer.

(f) Items that are required to be sterile and come into contact with blood should be disposable whenever possible.

[40 FR 53532, Nov. 18, 1975, as amended at 59 FR 23636, May 6, 1994]

**Subpart E [Reserved]**

**Subpart F—Production and Process Controls**

**§ 606.100 Standard operating procedures.**

(a) In all instances, except clinical investigations, standard operating procedures shall comply with published additional standards in part 640 of this chapter for the products being processed; except that, references in part 640 relating to licenses, licensed establishments and submission of material or data to or approval by the Director, Center for Biologics Evaluation and Research, are not applicable to establishments not subject to licensure under section 351 of the Public Health Service Act.

(b) Written standard operating procedures shall be maintained and shall include all steps to be followed in the collection, processing, compatibility testing, storage, and distribution of blood and blood components for transfusion and further manufacturing purposes. Such procedures shall be available to the personnel for use in the areas where the procedures are performed. The written standard operating procedures shall include, but are not limited to, descriptions of the following, when applicable:

(1) Criteria used to determine donor suitability, including acceptable medical history criteria.

(2) Methods of performing donor qualifying tests and measurements, including minimum and maximum values for a test or procedure when a factor in determining acceptability.

(3) Solutions and methods used to prepare the site of phlebotomy to give maximum assurance of a sterile container of blood.

(4) Method of accurately relating the product(s) to the donor.

(5) Blood collection procedure, including in-process precautions taken to measure accurately the quantity of blood removed from the donor.

(6) Methods of component preparation, including any time restrictions for specific steps in processing.

(7) All tests and repeat tests performed on blood and blood components during manufacturing.

(8) Pretransfusion testing, where applicable, including precautions to be taken to identify accurately the recipient blood samples and crossmatched donor units.

(9) Procedures for investigating adverse donor and recipient reactions.

(10) Storage temperatures and methods of controlling storage temperatures for all blood products and reagents as prescribed in §§ 600.15 and 610.53 of this chapter.

(11) Length of expiration dates, if any, assigned for all final products as prescribed in § 610.53 of this chapter.

(12) Criteria for determining whether returned blood is suitable for reissue.

(13) Procedures used for relating a unit of blood or blood component from the donor to its final disposition.

(14) Quality control procedures for supplies and reagents employed in blood collection, processing and pretransfusion testing.

(15) Schedules and procedures for equipment maintenance and calibration.

(16) Labeling procedures, including safeguards to avoid labeling mixups.

(17) Procedures of plasmapheresis, plateletpheresis, and leukapheresis, if performed, including precautions to be taken to ensure reinfusion of a donor's own cells.

(18) Procedures for preparing recovered plasma, if performed, including details of separation, pooling, labeling, storage, and distribution.

(19) Procedures in accordance with § 610.46 of this chapter to look at prior donations of Whole Blood, blood components, Source Plasma and Source Leukocytes from a donor who has donated blood and subsequently tests repeatedly reactive for antibody to human immunodeficiency virus (HIV) or otherwise is determined to be unsuitable when tested in accordance with § 610.45 of this chapter. Procedures to quarantine in-house Whole Blood, blood components, Source Plasma and Source Leukocytes intended for further manufacture into injectable products that were obtained from such donors; procedures to notify consignees regarding the need to quarantine such products; procedures to determine the suitability for release of such products from quarantine; procedures to notify consignees of Whole Blood, blood components, Source Plasma and Source Leukocytes from such donors of the results of the antibody testing of such donors; and procedures in accordance with § 610.47 of this chapter to notify attending physicians so that transfusion recipients are informed that they may have received Whole Blood and, blood components at increased risk for transmitting human immunodeficiency virus.

(20) Procedures for donor deferral as prescribed in § 610.41 of this chapter; and procedures for donor notification and autologous donor referring physician notification, including procedures for the appropriate followup if the initial attempt at notification fails, as prescribed in § 630.6 of this chapter.

(c) All records pertinent to the lot or unit maintained pursuant to these regulations shall be reviewed before the release or distribution of a lot or unit of final product. The review or portions of the review may be performed at appropriate periods during or after blood collecting, processing, compatibility testing and storing. A thorough investigation, including the conclusions and followup, of any unexplained discrepancy or the failure of a lot or unit to meet any of its specifications shall be made and recorded.

**§ 606.110**

**21 CFR Ch. I (4–1–07 Edition)**

(d) In addition to the requirements of this subpart and in conformity with this section, any facility may utilize current standard operating procedures such as the manuals of the organizations, as long as such specific procedures are consistent with, and at least as stringent as, the requirements contained in this part.

(1) American Association of Blood Banks.

(2) American National Red Cross.

(3) Other organizations or individual blood banks, subject to approval by the Director, Center for Biologics Evaluation and Research.

[40 FR 53532, Nov. 18, 1975, as amended at 49 FR 23833, June 8, 1984; 55 FR 11013, Mar. 26, 1990; 61 FR 47422, Sept. 9, 1996; 64 FR 45370, Aug. 19, 1999; 66 FR 31176, June 11, 2001]

**§ 606.110 Plateletpheresis, leukapheresis, and plasmapheresis.**

(a) The use of plateletpheresis and leukapheresis procedures to obtain a product for a specific recipient may be at variance with the additional standards for specific products prescribed in this part provided that: (1) A physician has determined that the recipient must be transfused with the leukocytes or platelets from a specific donor, and (2) the procedure is performed under the supervision of a qualified licensed physician who is aware of the health status of the donor, and the physician has certified in writing that the donor's health permits plateletpheresis or leukapheresis.

(b) Plasmapheresis of donors who do not meet the donor requirements of §§ 640.63, 640.64 and 640.65 of this chapter for the collection of plasma containing rare antibodies shall be permitted only with the prior approval of the Director, Center for Biologics Evaluation and Research.

[40 FR 53532, Nov. 18, 1975, as amended at 49 FR 23833, June 8, 1984; 55 FR 11013, Mar. 26, 1990]

**Subpart G—Finished Product Control**

**§ 606.120 Labeling, general requirements.**

(a) Labeling operations shall be separated physically or spatially from

other operations in a manner adequate to prevent mixups.

(b) The labeling operation shall include the following labeling controls:

(1) Labels shall be held upon receipt, pending review and proofing against an approved final copy, to ensure accuracy regarding identity, content, and conformity with the approved copy.

(2) Each type of label representing different products shall be stored and maintained in a manner to prevent mixups, and stocks of obsolete labels shall be destroyed.

(3) All necessary checks in labeling procedures shall be utilized to prevent errors in translating test results to container labels.

(c) All labeling shall be clear and legible.

[50 FR 35469, Aug. 30, 1985]

**§ 606.121 Container label.**

(a) The container label requirements are designed to facilitate the use of a uniform container label for blood and blood components (except Source Plasma) by all blood establishments.

(b) The label provided by the collecting facility and the initial processing facility shall not be removed, altered, or obscured, except that the label may be altered to indicate the proper name and other information required to identify accurately the contents of a container after blood components have been prepared.

(c) The container label shall include the following information, as well as other specialized information as required in this section for specific products:

(1) The proper name of the product in a prominent position, and modifier(s), if appropriate.

(2) The name, address, registration number, and, if a licensed product, the license number of each manufacturer.

(3) The donor, pool, or lot number relating the unit to the donor.

(4) The expiration date, including the day, month, and year, and, if the dating period for the product is 72 hours or less, the hour of expiration.

(5) If the product is intended for transfusion, the appropriate donor classification statement, i.e., "paid donor" or "volunteer donor", in no less