

Food and Drug Administration, HHS

§ 606.3

reporting applicable to all approved biological products.

§ 601.94 Promotional materials.

For biological products being considered for approval under this subpart, unless otherwise informed by the agency, applicants must submit to the agency for consideration during the preapproval review period copies of all promotional materials, including promotional labeling as well as advertisements, intended for dissemination or publication within 120 days following marketing approval. After 120 days following marketing approval, unless otherwise informed by the agency, the applicant must submit promotional materials at least 30 days prior to the intended time of initial dissemination of the labeling or initial publication of the advertisement.

§ 601.95 Termination of requirements.

If FDA determines after approval under this subpart that the requirements established in §§ 601.91(b)(2), 601.92, and 601.93 are no longer necessary for the safe and effective use of a biological product, FDA will so notify the applicant. Ordinarily, for biological products approved under § 601.91, these requirements will no longer apply when FDA determines that the postmarketing study verifies and describes the biological product's clinical benefit. For biological products approved under § 601.91, the restrictions would no longer apply when FDA determines that safe use of the biological product can be ensured through appropriate labeling. FDA also retains the discretion to remove specific postapproval requirements upon review of a petition submitted by the sponsor in accordance with § 10.30 of this chapter.

PART 606—CURRENT GOOD MANUFACTURING PRACTICE FOR BLOOD AND BLOOD COMPONENTS

Subpart A—General Provisions

Sec.
606.3 Definitions.

Subpart B—Organization and Personnel

606.20 Personnel.

Subpart C—Plant and Facilities

606.40 Facilities.

Subpart D—Equipment

606.60 Equipment.

606.65 Supplies and reagents.

Subpart E [Reserved]

Subpart F—Production and Process Controls

606.100 Standard operating procedures.

606.110 Plateletpheresis, leukapheresis, and plasmapheresis.

Subpart G—Finished Product Control

606.120 Labeling, general requirements.

606.121 Container label.

606.122 Instruction circular.

Subpart H—Laboratory Controls

606.140 Laboratory controls.

606.151 Compatibility testing.

Subpart I—Records and Reports

606.160 Records.

606.165 Distribution and receipt; procedures and records.

606.170 Adverse reaction file.

606.171 Reporting of product deviations by licensed manufacturers, unlicensed registered blood establishments, and transfusion services.

AUTHORITY: 21 U.S.C. 321, 331, 351, 352, 355, 360, 360j, 371, 374; 42 U.S.C. 216, 262, 263a, 264.

SOURCE: 40 FR 53532, Nov. 18, 1975, unless otherwise noted.

Subpart A—General Provisions

§ 606.3 Definitions.

As used in this part:

(a) *Blood* means whole blood collected from a single donor and processed either for transfusion or further manufacturing.

(b) *Unit* means the volume of blood or one of its components in a suitable volume of anticoagulant obtained from a single collection of blood from one donor.

(c) *Component* means that part of a single-donor's blood separated by physical or mechanical means.

§ 606.20

(d) *Plasma for further manufacturing* means that liquid portion of blood separated and used as material to prepare another product.

(e) *Plasmapheresis* means the procedure in which blood is removed from the donor, the plasma is separated from the formed elements and at least the red blood cells are returned to the donor.

(f) *Plateletpheresis* means the procedure in which blood is removed from a donor, a platelet concentrate is separated, and the remaining formed elements are returned to the donor along with a portion of the residual plasma.

(g) *Leukapheresis* means the procedure in which blood is removed from the donor, a leukocyte concentrate is separated, and the remaining formed elements and residual plasma are returned to the donor.

(h) *Facilities* means any area used for the collection, processing, compatibility testing, storage or distribution of blood and blood components.

(i) *Processing* means any procedure employed after collection and before compatibility testing of blood and includes the identification of a unit of donor blood, the preparation of components from such unit of donor blood, serological testing, labeling and associated recordkeeping.

(j) *Compatibility testing* means the procedures performed to establish the matching of a donor's blood or blood components with that of a potential recipient.

(k) *Distributed* means:

(1) The blood or blood components have left the control of the licensed manufacturer, unlicensed registered blood establishment, or transfusion service; or

(2) The licensed manufacturer has provided Source Plasma or any other blood component for use in the manufacture of a licensed biological product.

(l) *Control* means having responsibility for maintaining the continued safety, purity, and potency of the product and for compliance with applicable product and establishment standards,

21 CFR Ch. I (4-1-07 Edition)

and for compliance with current good manufacturing practices.

[40 FR 53532, Nov. 18, 1975, as amended at 64 FR 45370, Aug. 19, 1999; 65 FR 66635, Nov. 7, 2000; 66 FR 1835, Jan. 10, 2001; 66 FR 40889, Aug. 6, 2001]

Subpart B—Organization and Personnel

§ 606.20 Personnel.

(a) [Reserved]

(b) The personnel responsible for the collection, processing, compatibility testing, storage or distribution of blood or blood components shall be adequate in number, educational background, training and experience, including professional training as necessary, or combination thereof, to assure competent performance of their assigned functions, and to ensure that the final product has the safety, purity, potency, identity and effectiveness it purports or is represented to possess. All personnel shall have capabilities commensurate with their assigned functions, a thorough understanding of the procedures or control operations they perform, the necessary training or experience, and adequate information concerning the application of pertinent provisions of this part to their respective functions.

(c) Persons whose presence can adversely affect the safety and purity of the products shall be excluded from areas where the collection, processing, compatibility testing, storage or distribution of blood or blood components is conducted.

[40 FR 53532, Nov. 18, 1975, as amended at 49 FR 23833, June 8, 1984; 55 FR 11014, Mar. 26, 1990; 62 FR 53538, Oct. 15, 1997]

Subpart C—Plant and Facilities

§ 606.40 Facilities.

Facilities shall be maintained in a clean and orderly manner, and shall be of suitable size, construction and location to facilitate adequate cleaning, maintenance and proper operations. The facilities shall:

(a) Provide adequate space for the following when applicable: