

Food and Drug Administration, HHS

§ 660.34

the manufacturer's biologics license application.

[52 FR 37450, Oct. 7, 1987, as amended at 64 FR 56454, Oct. 20, 1999]

§ 660.31 Suitability of the donor.

Donors of peripheral blood for Reagent Red Blood Cells shall meet the criteria for donor suitability under § 640.3 of this chapter, except that paragraphs (b)(5) and (6), (d), and (e) of § 640.3 shall not apply.

§ 660.32 Collection of source material.

Blood for Reagent Red Blood Cells from donors of peripheral blood shall be collected as prescribed under § 640.4 of this chapter, except that paragraphs (c), (d), (g), and (h) of § 640.4 shall not apply.

§ 660.33 Testing of source material.

Except as provided in this section, a sample of each blood incorporated into the Reagent Red Blood Cell product shall be individually tested, with no fewer than two donor sources of each antibody specificity employed, to confirm the identification of all blood group antigens specified in the labeling as present or absent. The manufacturer shall perform at least one of the required tests for each factor. The Reagent Red Blood Cell product may be tested with a single donor source of antibody specificity if only one source of antibody is available, and the Director, Center for Biologics Evaluation and Research, has approved the use of a single donor source of antiserum. Each of these tests shall be conducted and interpreted independently, and any discrepancy between the results of these two tests shall be resolved by testing with at least one additional antiserum before concluding that the antigen is present or absent. Where fewer than three donor sources of an antibody specificity are available, test discrepancies shall be resolved in accordance with the manufacturer's biologics license application. Group O Reagent Red Blood Cells used in the detection or identification of unexpected antibodies shall include at least the following common antigens in each lot of the product: D, C, E, \bar{c} , e, K, \bar{k} , Fy^a,

Fy^b, Jk^a, Jk^b, Le^a, Le^b, P₁, M, N, S, and \bar{s} .

[52 FR 37450, Oct. 7, 1987, as amended at 55 FR 11013, Mar. 26, 1990; 64 FR 56454, Oct. 20, 1999]

§ 660.34 Processing.

(a) *Processing method.* The processing method shall be one that has been shown to yield consistently a product that is capable of detecting, throughout the dating period, alloantibodies corresponding to all required blood group antigens specified in the labeling as present.

(b) *Products prepared from pooled red blood cells.* If the product is recommended for the detection of unexpected antibodies, the pool shall be prepared by combining equal amounts of cells from no more than two donors. Umbilical cord cells are exempt from this requirement. Pooled cells shall not be recommended for pretransfusion tests, done in lieu of a major cross-match, to detect unexpected antibodies in patients' samples.

(c) *Absence of antibodies.* Each lot of final product shall be free of demonstrable antibodies, including anti-A and anti-B, unless the package insert and container label include instructions to wash the cells before use. The final product shall also be direct antiglobulin test negative when tested with polyspecific anti-human globulin.

(d) *Final container.* The final containers used for each lot of product shall be clean and shall permit observation of the contents for hemolysis or a change in color. The final container label, container cap, and dropper bulb of a Reagent Red Blood Cell product may be color-coded with a visual match to a specific color approved by the Director, Center for Biologics Evaluation and Research.

(e) *Date of manufacture.* The date of manufacture of the product shall be the date that the blood is withdrawn from the donor or obtained from umbilical cords. The period during which the reagent red blood cell source material is kept by the manufacturer in storage in a frozen state at -65°C or colder is excluded from the dating period. If the product consists of red blood cells from two or more donors, the date of manufacture of the final product shall be the date of withdrawal of blood from the