

### § 660.43

standards of purity and quality and shall be effectively segregated and otherwise manufactured in a manner that will reduce the risk of contaminating the product and other biological products. Ancillary reagents and materials accompanying the product, which are used in the performance of the test as described by the manufacturer's recommended test procedures, shall have been shown not to affect adversely the product within the prescribed dating period.

(c) *Final container.* A final container shall be sufficiently transparent to permit visual inspection of the contents for presence of particulate matter and increased turbidity. The effectiveness of the contents of a final container shall be maintained throughout its dating period.

(d) *Date of manufacture.* The date of manufacture of Hepatitis B Surface Antigen that has been iodinated with radioactive iodine (<sup>125</sup>I) shall be the day of labeling the antibody with the radionuclide.

[44 FR 36382, June 22, 1979, as amended at 49 FR 1685, Jan. 13, 1984]

### § 660.43 Potency test.

To be satisfactory for release, each filling of Hepatitis B Surface Antigen shall be tested against the Reference Hepatitis B Antiserum Panel and shall be sufficiently potent to be able to detect the antibody in the appropriate sera of the reference panel by all test methods recommended by the manufacturer in the package insert.

### § 660.44 Specificity.

Each filling of the product shall be specific for Hepatitis B Surface Antigen as determined by specificity tests found acceptable to the Director, Center for Biologics Evaluation and Research.

[44 FR 36382, June 22, 1979, as amended at 49 FR 23834, June 8, 1984; 55 FR 11013, Mar. 26, 1990]

### § 660.45 Labeling.

In addition to the requirements of §§ 610.60, 610.61, and 809.10 of this chapter, the labeling shall bear the following:

### 21 CFR Ch. I (4-1-08 Edition)

(a) The "d and y" antigen subtype and the source of the product to follow immediately the proper name on both the final container label and the package label. If the product is intended to identify antibodies to the "r and w" antigen subtype, the antigen subtype designation shall include the "r and w" antigen subtype.

(b) The name of the test method(s) recommended for use of the product on the package label and on the final container label, when capable of bearing a full label (see § 610.60(a) of this chapter).

(c) A warning on the package label and on the final container label stating that the product is capable of transmitting hepatitis and should be handled accordingly.

(d) The package shall include a package insert providing (1) detailed instructions for use, (2) an adequate description of all recommended test methods, and (3) warnings as to possible hazards, including hepatitis transmitted in handling the product and any ancillary reagents and materials accompanying the product.

### § 660.46 Samples; protocols; official release.

(a) *Samples.* (1) For the purposes of this section, a sample of product not iodinated with <sup>125</sup>I means a sample from each filling of each lot packaged as for distribution, including all ancillary reagents and materials; and a sample of product iodinated with <sup>125</sup>I or unlyophilized HBsAg-coated red blood cells means a sample from each lot of diagnostic test kits in a finished package, including all ancillary reagents and materials.

(2) Unless the Director, Center for Biologics Evaluation and Research, determines that the reliability and consistency of the finished product can be assured with a smaller quantity of sample or no sample and specifically reduces or eliminates the required quantity of sample, each manufacturer shall submit the following samples to the Director, Center for Biologics Evaluation and Research (see mailing addresses in § 600.2 of this chapter), within 5 working days after the manufacturer has satisfactorily completed all tests on the samples: